

Supplement

Manual Title: 80PK-3A I. S. Supplement Issue: **1**
Print Date: Sept. 1989 Issue Date: **3/10**
Revision/Date: 4, 4/02 Page Count: **1**

This supplement contains information
necessary to ensure the accuracy of the
above Instruction Sheet.

© 2010 Fluke Corporation. All rights reserved. Printed in U.S.A.

FLUKE®

Change #1, 54379

Under **Specifications**, change the following:

From: **Junction Accuracy:** (With respect to ANSI MC96.1):
 $\pm 2.2\text{ }^{\circ}\text{C}$ (3.96 $^{\circ}\text{F}$) over the range of 0 $^{\circ}\text{C}$ to 260 $^{\circ}\text{C}$
(32 $^{\circ}\text{F}$ to 500 $^{\circ}\text{F}$)

To: **Accuracy:** (With respect to ANSI MC96.1): $\pm 4.0\text{ }^{\circ}\text{C}$
(7.2 $^{\circ}\text{F}$) over the range of 0 $^{\circ}\text{C}$ to 260 $^{\circ}\text{C}$ (32 $^{\circ}\text{F}$ to 500 $^{\circ}\text{F}$)